Training on a virtual reality cricothyroidotomy simulator improves skills and transfers to a simulated procedure

Ganesh Sankaranarayanan ¹, Coleman A Odlozil,¹ Salman S Hasan,¹ Rehman Shabbir,¹ Di Qi,² Melih Turkseven,² Suvranu De,² Geoffrey Funk,¹ Rebecca J Weddle¹

► Additional supplemental material is published online only. To view, please visit the journal online (<http://dx.doi.org/10.1136/tsaco-2021-000826>).

¹Baylor University Medical Center at Dallas, Dallas, Texas, USA

²Mechanical, Aerospace and Nuclear Engineering, Rensselaer Polytechnic Institute, Troy, NY, USA

Correspondence to

Dr Ganesh Sankaranarayanan; ganesh.sankaranarayanan@utsouthwestern.edu

Received 24 August 2021

Accepted 9 January 2022

ABSTRACT

Objective The virtual airway skills trainer (VAST) is a virtual reality simulator for training in cricothyroidotomy (CCT). The goal of the study is to test the effectiveness of training and transfer of skills of the VAST-CCT.

Methods Two groups, control (no training) and simulation (2 weeks of proficiency-based training), participated in this study. Subjects in the control condition did not receive any training on the task whereas those in the simulation received a proficiency-based training on the task during a period of 2 weeks. Two weeks post-training, both groups performed CCT on the TraumaMan to demonstrate the transfer of skills.

Results A total of (n=20) subjects participated in the study. The simulation group performed better than the control group at both the post-test (p<0.001) and retention test (p<0.001) on the simulator. The cumulative sum analysis showed that all subjects in the simulation group reached proficiency with acceptable failure rate within the 2 weeks of training. On the transfer test, the simulation group performed better on skin cut (p<0.001), intubation (p<0.001) and total score (p<0.001) than the control group.

Conclusions The VAST-CCT is effective in training and skills transfer for the CCT procedure.

Level of evidence Not applicable. Simulator validation study.

INTRODUCTION

Cricothyroidotomy (CCT) is an emergency life-saving procedure for patients who cannot be intubated by conventional means and would otherwise face impending death.¹ Since CCTs are performed infrequently, with incidence of both prehospital and in-hospital CCTs requiring secure airway, ranging from 0% to 18.5%,¹ training on live patients is limited. Currently, learning and maintaining the skill is done by practicing on cadavers, animal models, manikins or small benchtop models.^{2–12} Due to ethical issues in using animals, 99% of the advanced trauma life support (ATLS) courses in the USA and all of the courses in Canada now use the American College of Surgeons (ACS)-approved manikin-based simulators.¹³ The main limitations of manikin-based simulators are in providing a high-fidelity simulation with variations that are routinely seen in clinical cases.

To overcome these challenges, we developed the virtual airway skills trainer (VAST), a virtual reality simulator for training in airway procedures. The VAST-CCT, developed for training in the CCT procedure, is a validated simulator that has been previously shown to differentiate performance of surgeons with more than five actual cricothyrotomies.¹⁴

The goal of this work is to assess the ability of the VAST-CCT simulator to train novice subjects in the CCT procedure and transfer knowledge to a simulated procedure. It is hypothesized that subjects trained in the VAST-CCT procedure will demonstrate improvement in skill and positive transfer of learning in the simulated clinical environment compared with control group with no prescribed training.

MATERIALS AND METHODS

Study design and procedure

The study was a between-subjects design with medical school student volunteers randomized into two groups: control and simulation.

Subjects

Twenty (n=20) (13 men and 7 women) second year and third year Texas A&M College of Medicine students were recruited at the Baylor University Medical Center in Dallas for this study. The subjects were randomly assigned to either control (n=10) or simulation (n=10) groups. The gender distribution in our randomly assigned groups was as follows: control (male=6, female=4) and the simulation (male=7, female=3).

Pretest

To assess the baseline performance (pretest), both groups performed the task once on the VAST-CCT simulator at the beginning of the study. The subjects were first introduced to the VAST-CCT simulator before completing the task once for their baseline assessment.

Training

Subjects who were assigned to the simulation group then practiced the task one session a day, 5 days/week, for two consecutive weeks, for a total of 10 training sessions. At the beginning of the first training session, they were also asked to watch an

© Author(s) (or their employer(s)) 2022. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

To cite: Sankaranarayanan G, Odlozil CA, Hasan SS, et al. *Trauma Surg Acute Care Open* 2022;**7**:e000826.

online supplemental video of an expert performing the CCT on the simulator. During the training sessions, subjects were allowed unlimited attempts until they demonstrated proficiency-level performance for two consecutive repetitions. The subjects in the simulation group were monitored in person during the training session by the experimenters led by the study lead, who were trained in all aspects of this procedure to provide feedback by our trauma and critical care surgeons. After each trial, the subjects' performance scores on landmark identification, skin and membrane incision and intubation were disclosed and feedback was provided (eg, showing proper landmark position, showing them the start-end position as well as direction of the incision and whether endotracheal tube is placed too deep or shallow). Further instruction was given to the subjects if they appeared to be struggling with the task. Each session was no longer than an hour. Subjects in the control group did not participate in the training sessions.

Proficiency-based training

A proficiency-based model¹⁵ was used to train the subjects in the simulation group. The learning rate and plateau levels of subjects vary within a group, and a preset or arbitrary end criterion may not ensure that all subjects reach the level of proficiency. However, proficiency criteria derived from expert performance have been successfully used in training subjects in surgical skills.^{15 16} For surgical motor tasks, the mean performance value of expert performance has been used as a proficiency metric.¹⁶ In our study, two expert board-certified trauma surgeons were asked to perform the CCT task five times on the VAST-CCT after becoming familiarized on the simulator. The automatically computed score for the five consecutive repetitions was then used to calculate the proficiency score for the simulation group. Since our subject population are medical students who are completely novice to this procedure, we were interested in assessing the elevation of their performance from novice level and kept the proficiency score as mean—SD to be an achievable target. Demonstration of performance at this proficiency level for two or more consecutive repetitions was marked as the endpoint for each day of training.

Post-test and retention test

At the end of the 10 training sessions, all subjects performed the task once on the VAST-CCT simulator (post-test). Two weeks after the post-test, all subjects performed the task again to determine their skill retention (retention test). The washout period of 2 weeks was chosen based on our previous surgical skills training with medical students.^{17 18}

Transfer test

Immediately after the retention test, subjects performed CCT on a TraumaMan (Simulab, Seattle, Washington) manikin to demonstrate their transfer skills. We chose TraumaMan because it is an ACS-approved training platform for the ATLS course. Transfer of skills to the clinical environment is desirable but risky as our subject population are medical students and cannot yet operate on human patients. Cadavers have limitations of being expensive, single use, and without the same feel as live tissue. Each subject repeated the task twice on the manikin and the entire procedure was videotaped. The completion time, skin and membrane cut lengths, and intubation length were recorded and a score was computed using the same procedure in our simulator. Additionally, two expert trauma and critical care surgeons rated the performance of the subjects from the

Figure 1 The virtual airway skills trainer for cricothyroidotomy (VAST-CCT) simulator showing the visual and haptic interface.

recorded videos. The videos were double blinded (both groups and repetitions). We derived a subjective assessment tool based on the Objective Structured Assessment of Technical Skills¹⁹ using which the experts rated the performance of the subjects on a 5-point Likert scale for preparation, respect for tissue, time and motion, instrument handling, knowledge of procedure, and overall performance. The experts also rated the performance of each of the subjects using a custom checklist as the available validated tool was for rapid four-step technique.²⁰ The checklist had seven items that were scored either 0 (does not perform the task), 1 (performs the task inadequately) or 2 (performs the task adequately).

Simulator

The VAST-CCT is a validated simulator that was developed for training in critical airway skills.¹⁴ It consists of a specialized palpation interface for landmark identification and a haptic feedback device for force feedback while performing the task (figure 1). A computer monitor is used to display the virtual anatomic models and the tools used in the procedure. The simulator is designed to train in four key tasks of the CCT procedure: (1) landmark identification, (2) skin incision, (3) cricothyroid membrane incision, and (4) intubation. During the landmark identification task, the users manually palpate the interface to identify the cricoid and thyroid cartilages. An array of force sensors mounted underneath the palpation interface is used to triangulate the position of the finger for display on the monitor. Once the landmarks are identified, users can hold on to the handle of the haptic device (Geomagic Touch, 3D Systems, Rock Hill, South Carolina) to choose appropriate tools to complete the other three tasks. The simulator automatically records performance metrics for assessment. A more detailed description of the simulation of the procedure in the VAST-CCT is described in our construct validation article—the VAST-CCT simulator.¹⁴

Statistical analysis

The performance at pretest, post-test, and retention test results on the VAST-CCT simulator was analyzed using a separate two-way mixed factorial analysis of variance (ANOVA) with 'Test' (pre, post, and retention) as the within-subjects factor and the 'Group' (control and simulation) as the between-subjects factor for both completion time and total score. Interaction between the two (Test*Group) was also analyzed. For further analyzing the significance a simple effects test with Bonferroni correction was used. The cumulative sum (CUSUM) method was used to analyze the learning of the simulation group during the training.²¹

The CUSUM is a useful statistical tool for quality control of sequential process.²¹ The CUSUM chart is plotted by first setting the acceptable (p_o) and unacceptable failure rates (p_i), probability of type I (α) and type II (β) errors. Based on these values, the upper (h_1) and lower (h_0) decision limits and a constant s were calculated. At each trial, if an outcome is a success, the constant s is subtracted from the previous CUSUM. If the outcome is failure then $1-s$ is added to the previous CUSUM. When successive trials are successful the CUSUM score will decrease (negative slope) whereas in the case of failure it will increase (positive slope), which can be visualized by plotting the CUSUM. By plotting the CUSUM over time, the time at which the process reaches acceptable failure rate can be determined. It has been widely used for the analysis of the learning curve in many medical procedures.^{17 18 22–31}

The experts' ratings on the transfer tests were tested for inter-rater reliability by calculating the Cronbach's alpha and then analyzed using a non-parametric Mann-Whitney U test for significance between the groups.

RESULTS

Pretest, post-test and retention test results

The subject's performance on the VAST-CCT, including completion time (figure 2A,B) and total score (figure 2C,D), for both the control and simulation groups, is shown in figure 2. The results from the two-way mixed factorial ANOVA are summarized in table 1.

Figure 2 Performance at pretest, post-test and retention test for (A) completion time—control group, (B) completion time—simulation group, (C) total score—control group, (D) total score—simulation group.

Table 1 Two-way mixed ANOVA results for completion time and total score

Factors	F score	P value	Partial eta squared	Observed power
Completion time in seconds				
<i>Within-subjects effects</i>				
Test	$F(2,36)=38.63$	<0.001	0.682	1.0
Test*Group	$F(2,36)=12.36$	<0.001	0.407	0.993
<i>Between-subjects effects</i>				
Group	$F(1,18)=9.25$	0.007	0.34	0.82
Total score				
<i>Within-subjects effects</i>				
Test	$F(2,36)=0.89$	0.418	0.047	0.19
Test*Group	$F(2,36)=898.2$	<0.001	0.588	1.0
<i>Between-subjects effects</i>				
Group	$F(1,18)=1269$	<0.001	0.556	0.99

ANOVA, analysis of variance.

Analysis of the completion time showed a significant main effect of test ($F(2,36)=38.63$, $p<0.001$, $\eta^2=0.682$) and interaction (Test*Group) ($F(2,36)=12.36$, $p<0.001$, $\eta^2=0.407$). This shows that there was a significant change in completion time between the pretest, post-test, and retention test. There was also a significant difference in performance between the groups ($F(1,18)=9.25$, $p=0.007$, $\eta^2=0.34$). To analyze the significant results in completion time further, a simple effects test with Bonferroni correction for multiple comparisons was computed. At pretest, both groups' baseline completion time was not significantly different (162.7 vs. 193.2, $p=0.32$), but the simulation group completed the task faster than the control group at the post-test (128.8 vs. 42.1, $p<0.001$) and retention test (112.1 vs. 47.3, $p<0.001$). The control group's mean completion time did not show significant differences between pretest and post-test (162.7 vs. 128.8, $p=0.42$), pretest and retention test (162.7 vs. 112.1, $p=0.064$) and post-test and retention test (128.8 vs. 112.1, $p=0.122$) whereas the simulation group showed improvement in completion time between pretest and post-test (193.2 vs. 42.1, $p<0.001$), pretest and retention test (193.2 vs. 47.3, $p<0.001$) and skill retention between post-test and retention test (42.1 vs. 47.3, $p=1.0$).

For the total score, the main effect of test was not significant ($F(2,36)=0.89$, $p=0.418$, $\eta^2=0.047$) but there was a significant interaction (Test*Group) ($F(2,36)=898.2$, $p<0.001$, $\eta^2=0.588$). There was also a significant difference in total score between the two groups ($F(1,18)=1269.6$, $p<0.001$, $\eta^2=0.556$). Test of simple effects showed that at pretest there was no significant difference in total score between the control and simulation groups (28.0 vs. 22.2, $p=0.1$). During the post-test, the simulation group performed significantly better than the control group as expected (17.4 vs. 37.4, $p<0.001$) and during the retention test, the simulation group continued to perform better than the control group indicating skill retention (20.4 vs. 33.8, $p<0.001$). As the main effect was insignificant, no further simple effects tests were conducted.

Learning curve analysis

Proficiency score

The average score from the experts' ($n=2$) performance on the VAST-CCT was 34 ± 4 . The proficiency score for our study was set as 1 SD below the mean: 30 ($34-4$).

Figure 3 Cumulative sum (CUSUM) chart of the individual subjects in the simulation group. (Note: subjects were given an identifier based on randomization order generated for a total of 25 subjects to account for attrition.)

CUSUM analysis

For the analysis of the CCT data, with the proficiency score of 30, acceptable failure rate p_o of 20% (0.2), $p_1=0.4$, $\alpha=0.5$, $\beta=0.2$, the upper and lower decision limits and the constant s , computed were $h_0 = -1.58$, $h_1 = 2.82$ and $s=0.3$, respectively.

Figure 3 shows the CUSUM plot of all subjects in the simulation. Based on the proficiency score of 30, all subjects (MS24 at the 9th trial, MS14 at the 12th trial, MS18 at the 13th trial, MS12 at the 23rd trial, MS20 at the 27th trial, MS1 and MS19 at the 33rd trial and MS7, MS8 and MS9 at the 67th trial) reached the level of acceptable 20% failure rate indicated by their individual CUSUM score crossing the lower decision limit (h_0).

Transfer test results

Figure 4 shows a subject performing the transfer test on the TraumaMan manikin. The video of a subject performing the transfer task is provided as an online supplemental material 1. In all of our analyses of the transfer task, we excluded two subjects in the control group who had to perform the transfer task on a SimMan manikin due to unavailability of the TraumaMan manikin at the time of the study.

Figure 4 Subject performing the cricothyroidotomy (CCT) on the TraumaMan manikin.

Figure 5 Computed metrics from the transfer test: (A) skin cut score, (B) intubation, (C) total score.

Analysis of the computed score

The metrics computed from the data, landmark identification, skin cut score, membrane cut score, intubation score, total score, and completion time were analyzed using a Mann-Whitney U test to assess the difference in performance between the control and the simulation group. The results from skin cut score (figure 5A) ($U=54$, $p<0.001$), intubation (figure 5B) ($U=6$, $p<0.001$) and total score (figure 5C) ($U=6.5$, $p<0.001$) showed that the simulation group performed better than the control group. The landmark ($U=144$, $p=0.78$), membrane cut score ($U=135$, $p=0.4$) and completion time ($U=135$, $p=0.4$) did not show a significant difference.

Expert rating assessment

The ratings of the two experts both on the global rating scale of performance (Cronbach's $\alpha=0.840$ and intraclass correlation coefficient (ICC)=0.724, $p<0.0001$) and task-specific checklist (Cronbach's $\alpha=0.988$ and ICC=0.976, $p<0.0001$) showed a high inter-rater agreement.

The Mann-Whitney U test results from the subject rating on the global rating scale showed none of the items in the rating scale were significant between the two groups: (1) preparation for procedure ($U=726$, $p=0.5$), (2) respect for tissue ($U=717$, $p=0.6$), (3) time and motion ($U=707$, $p=0.6$), (4) instrument handling ($U=683.5$, $p=0.4$), (5) flow of procedure ($U=747$, $p=0.8$), (6) knowledge of procedure ($U=738.5$, $p=0.8$), (7) overall performance ($U=636.5$, $p=0.1$), and (8) total score ($U=745$, $p=0.8$).

The Mann-Whitney U test results for the checklist score were: (1) identifies thyroid cartilage ($U=720$, $p=0.14$), (2) identifies cricoid cartilage ($U=760$, $p=1.0$), (3) makes a 4–6 cm skin incision between the cartilages ($U=714$, $p=0.5$), (4) dilates the skin incision ($U=752$, $p=0.8$), (5) makes an incision along the cricothyroid membrane ($U=680$, $p=0.03$), (6) dilates the incision in the cricothyroid membrane ($U=554$, $p=0.002$), (7) intubation ($U=668$, $p=0.2$), and (8) total ($U=702$, $p=0.5$). Both checklist item 5 'Makes an incision along the cricothyroid membrane' and item 6 'Dilates the incision in the cricothyroid membrane' showed simulation group performance was better than the control group.

DISCUSSION

Cricothyrotomies are one of the most infrequently performed procedures, yet they have one of the highest potentials to resuscitate a patient when all other attempts have failed. The availability of real clinical training for this procedure is rare due to early detection of a compromised airway. In a level 1 trauma center of 2741 trauma admissions (between April 2010 and February 2012), only four cricothyrotomies were performed among them for an incidence rate of 0.15%.³² While this is excellent for patient care, it creates a challenging situation to appropriately train physicians for a situation in which they have to perform a cricothyrotomy. Therefore, simulation-based training is critical

for learning and maintaining emergency airway management skills including CCT.^{33–35}

Manikin-based training (SimMan (Laerdal), TraumaMan (Simulab), etc) needs replacement of the skin after one or two attempts and cannot reproduce difficult airway situations. Cadavers can be useful for training in CCT as they can naturally present differences in anatomy unlike a manikin; however, they can be very expensive. It has been shown that cadaver-based training is superior in fidelity to training on manikins.¹² Moreover, all of these simulators also require a proctor to record data and score the performance.

The VAST-CCT simulator does not need replenishment of materials nor a proctor to time and score the performance as it is done automatically by the computer program. In our study, we established that the VAST-CCT can be effectively used to train the necessary skills to successfully perform the CCT procedure. Of the 20 medical students, the subjects randomized into the simulation group showed a significant decrease in completion time between pretest and post-test compared with the control group. Similarly, their scores also increased considerably between pretest and post-test indicating gain in proficiency. The control group scores were higher than the simulation group at the pretest because we conducted the baseline assessment after randomization and, coincidentally, the median performance score was higher than the simulation group at the pretest. At both post-test and retention test, the control group's performance decreased, indicating no significant learning of the CCT skills. These results are in line with expected gain in skill and proficiency through longitudinal training of novice subjects using a virtual reality simulator that has been observed in simulators developed to teach basic laparoscopic surgical tasks.^{17,18} In our study, by training the simulation group using a proficiency benchmark, the total number of trials that need to achieve proficiency in this task was reduced as opposed to a fixed number of trials (eg, 10 trials per day to a total of 100 trials).

The CUSUM analysis showed that the individual learning rate and the number of trials to reach proficiency varied significantly with one subject (MS24) which started to perform consistently above the proficiency criteria by the ninth trial as opposed to two subjects (MS8 and MS9) who took 67 trials to reach the consistency at the proficiency level in their performance. Since we trained the simulation group for only 10 sessions during 2 weeks, it is not possible to say whether these two subjects' performance continued this trend of consistent performance; nevertheless, within 2 weeks, all subjects performed at the set proficiency level.

To test the transfer of skills, we decided to use the TraumaMan (Simulab) as the standardized platform due to medical students not yet qualified to operate on a real patient and the convenience of providing similar anatomy and conditions for all the subjects as opposed to a cadaver or benchtop models. We replaced the neck, skin, and the cricothyroid membrane for each subject, which enabled us to measure the performance metrics for each subject. The simulation group's performance from the computed metrics was significantly higher than the control group for skin cut, intubation, and the total score, clearly demonstrating their proficiency in performing the CCT procedure. We speculate the reason there was no difference in landmark identification and membrane cut score was because the TraumaMan had prominent thyroid and cricoid cartilages that were not very difficult to identify and both groups had difficulty cutting the membrane as it was attached to the manikin using two sticky pads, which did not provide sufficient tension for cutting. Moreover, completion time between groups did not show significant results. We think

it is due to the difficulty both groups faced with cutting the skin as well as familiarization with the manikin.

We also assessed the performance using subjective ratings by two experts from the recorded videos. Our assessment tool showed excellent consistency and inter-rater reliability; however, the Likert scale assessment items were not useful in differentiating performance between the two groups. Since the transfer test was done at the end after the retention test, even the control group subjects had the opportunity to perform the task thrice on the simulator to understand the steps. Additionally, with the standard medical student training, they had baseline knowledge on how to use a scalpel and dilators. It underscores the need for development of assessment tools specific to CCT. Our task-specific checklist items did show difference in performance between the groups for making the incision along the cricothyroid membrane and dilation of the incision. These are important steps that the simulation group had practiced several times and it was not surprising that the simulation group's performance was superior. Due to the limitation of the manikin, as well as the necessity to score from recorded videos, other checklist items did not show significant results. As part of our future work, we will investigate on developing proper assessment tools for scoring the performance on the manikin that can then be used for simulation using cadavers or benchtop models.

In conclusion, our results demonstrate that the subjects in the simulation group had significant improvements in performance compared with the control group. Our results are due to both the efficacy of the VAST-CCT simulation trainer as well as the longitudinal training that subjects received. Others have had success with developing comparable models to today's training standard and have shown to perform similarly with no clear statistical difference between the groups.¹⁰ We think that the VAST-CCT will be a useful tool for teaching CCT to a novice trainee as well as a platform to maintain skills for experienced practitioners. A further study on subjects trained on the simulator and their performance on a real patient is needed to show predictive validity on all aspects of the simulator.

Acknowledgements The authors wish to acknowledge the funding from the National Heart, Lung, and Blood Institute.

Contributors All the authors have seen and approved this article and contributed equally to the submission. GS contributed to study design, data acquisition, analysis and drafting of the article and was the primary author of the article. CAO, SSH and RS contributed to data acquisition, analysis and drafting of the article. DQ and MT contributed to the development of the simulator and drafting of the article. GF and RJW contributed to study design, clinical feedback on the simulator, data acquisition, analysis and article drafting. SD contributed to study design, simulation development, article drafting and final approval before submission. SD is the guarantor.

Funding This study was funded by the National Heart, Lung, and Blood Institute (R01 HL119248).

Competing interests None declared.

Patient consent for publication Not applicable.

Ethics approval This study involves human participants and was approved by the Baylor Scott & White Research Institute Institutional Review Board (protocol number: 018-523). Participants gave informed consent to participate in the study before taking part.

Provenance and peer review Not commissioned; externally peer reviewed.

Data availability statement Data are available upon reasonable request. Data will be made available upon request.

Open access This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

ORCID iD

Ganesh Sankaranarayanan <http://orcid.org/0000-0003-1556-2797>

REFERENCES

- Langvad S, Hylmo PK, Nakstad AR, Vist GE, Sandberg M. Emergency cricothyrotomy - a systematic review. *Scand J Trauma Resusc Emerg Med* 2013;21:43.
- Varaday SS, Yentis SM, Clarke S. A homemade model for training in cricothyrotomy. *Anaesthesia* 2004;59:1012–5.
- John B, Suri I, Hillermann C, Mendonca C. Comparison of cricothyroidotomy on manikin vs. simulator: a randomised cross-over study. *Anaesthesia* 2007;62:1029–32.
- Surgical D, Care T. Simulation in trauma education: beyond ATLS. *Injury* 2014;45:817–8.
- Buonopane CE, Pasta V, Sottile D, Del Vecchio L, Maturo A, Merola R, Panunzi A, Urciuoli P, D'Orazi V. Cricothyrotomy performed with the Melker™ set or the QuickTrach™ kit: procedure times, learning curves and operators' preference. *G Chir* 2014;35:165–70.
- Melchioris J, Todsen T, Nilsson P, Kohl AP, Böttger M, Charabi B, Konge L, von Buchwald C. Self-directed simulation-based training of emergency cricothyroidotomy: a route to lifesaving skills. *Eur Arch Otorhinolaryngol* 2016;273:4623–8.
- Hart D, Rush R, Rule G, Clinton J, Beilman G, Anders S, Brown R, McNeil MA, Reihisen T, Chipman J, et al. Training and assessing critical airway, breathing, and hemorrhage control procedures for trauma care: live tissue versus synthetic models. *Acad Emerg Med* 2018;25:148–67.
- Katayama A, Nakazawa H, Tokumine J, Lefor AK, Watanabe K, Asao T, Yorozu T. A high-fidelity simulator for needle cricothyroidotomy training is not associated with increased proficiency compared with conventional simulators: a randomized controlled study. *Medicine* 2019;98:e14665.
- White L, Bly RA, D'Auria D, Aghdasi N, Bartell P, Cheng LL, Hannaford B. Cricothyrotomy simulator with computational skill assessment for procedural skill training in the developing world. *Otolaryngol Head Neck Surg* 2013;149:P60.
- Wray A, Khan F, Ray J, Rowe R, Boysen-Osborn M, Wiechmann W, Toohey S. Comparison of the bleeding cricothyrotomy model to SimMan for training students and residents emergency cricothyrotomy. *J Adv Med Educ Prof* 2019;7:144–8.
- Combes RD. A critical review of anaesthetised animal models and alternatives for military research, testing and training, with a focus on blast damage, haemorrhage and resuscitation. *Altern Lab Anim* 2013;41:385–415.
- Takayasu JK, Peak D, Stearns D. Cadaver-based training is superior to simulation training for cricothyrotomy and tube thoracostomy. *Intern Emerg Med* 2017;12:99–102.
- Gala SG, Crandall ML. Global collaboration to Modernize advanced trauma life support training. *J Surg Educ* 2019;76:487–96.
- Qi D, Petrusa E, Kruger U, Milef N, Abu-Nuwar MR, Haque M, Lim R, Jones DB, Turkseven M, Demirel D, et al. Surgeons with five or more actual Cricothyrotomies perform significantly better on a virtual reality simulator. *J Surg Res* 2020;252:247–54.
- Gallagher AG, Ritter EM, Champion H, Higgins G, Fried MP, Moses G, Smith CD, Satava RM. Virtual reality simulation for the operating room: proficiency-based training as a paradigm shift in surgical skills training. *Ann Surg* 2005;241:364–72.
- Ritter EM, Scott DJ. Design of a proficiency-based skills training curriculum for the fundamentals of laparoscopic surgery. *Surg Innov* 2007;14:107–12.
- Linsk AM, Monden KR, Sankaranarayanan G, Ahn W, Jones DB, De S, Schwaitzberg SD, Cao CGL. Validation of the VBLaST pattern cutting task: a learning curve study. *Surg Endosc* 2018;32:1990–2002.
- Zhang L, Sankaranarayanan G, Arikatla VS, Ahn W, Grosdemouge C, Rideout JM, Epstein SK, De S, Schwaitzberg SD, Jones DB, et al. Characterizing the learning curve of the VBLaST-PT(©) (Virtual Basic Laparoscopic Skill Trainer). *Surg Endosc* 2013;27:3603–15.
- Martin JA, Regehr G, Reznick R, MacRae H, Murnaghan J, Hutchison C, Brown M. Objective structured assessment of technical skill (OSATS) for surgical residents. *Br J Surg* 1997;84:273–8.
- Melchioris J, Todsen T, Nilsson P, Wennervaldt K, Charabi B, Böttger M, Konge L, von Buchwald C. Preparing for emergency: a valid, reliable assessment tool for emergency cricothyroidotomy skills. *Otolaryngol Head Neck Surg* 2015;152:260–5.
- Williams SM, Parry BR, Schlup MM. Quality control: an application of the cusum. *BMJ* 1992;304:6838:1359–61.
- Van Rij AM, McDonald JR, Pettigrew RA, Putterill MJ, Reddy CK, Wright JJ. Cusum as an aid to early assessment of the surgical trainee. *Br J Surg* 1995;82:1500–3.
- Pendlimari R, Holubar SD, Dozois EJ, Larson DW, Pemberton JH, Cima RR. Technical proficiency in hand-assisted laparoscopic colon and rectal surgery: determining how many cases are required to achieve mastery. *Arch Surg* 2012;147:317–22.
- Tekkis PP, Senagore AJ, Delaney CP, Fazio VW. Evaluation of the learning curve in laparoscopic colorectal surgery. *Ann Surg* 2005;242:83–91.
- Calsina L, Clará A, Vidal-Barraquer F. The use of the CUSUM chart method for surveillance of learning effects and quality of care in endovascular procedures. *Eur J Vasc Endovasc Surg* 2011;41:679–84.
- Bartlett A, Parry B. Cusum analysis of trends in operative selection and conversion rates for laparoscopic cholecystectomy. *ANZ J Surg* 2001;71:453–6.
- Naik VN, Devito I, Halpern SH. Cusum analysis is a useful tool to assess resident proficiency at insertion of labour epidurals. *Can J Anaesth* 2003;50:694–8.
- Biau DJ, Resche-Rigon M, Godiris-Petit G, Nizard RS, Porcher R. Quality control of surgical and interventional procedures: a review of the CUSUM. *Qual Saf Health Care* 2007;16:203–7.
- Okrainec A, Ferri LE, Feldman LS, Fried GM. Defining the learning curve in laparoscopic paraesophageal hernia repair: a CUSUM analysis. *Surg Endosc* 2011;25:1083–7.
- Fraser SA, Bergman S, Garzon J. Laparoscopic splenectomy: learning curve comparison between benign and malignant disease. *Surg Innov* 2012;19:27–32.
- Fraser SA, Feldman LS, Stanbridge D, Fried GM. Characterizing the learning curve for a basic laparoscopic drill. *Surg Endosc* 2005;19:1572–8.
- Schauer SG, Bellamy MA, Mabry RL, Bebaria VS. A comparison of the incidence of cricothyrotomy in the deployed setting to the emergency department at a level 1 military trauma center: a descriptive analysis. *Mil Med* 2015;180:60–3.
- Jayaraman V, Feeney JM, Brautigam RT, Burns KJ, Jacobs LM. The use of simulation procedural training to improve self-efficacy, knowledge, and skill to perform cricothyroidotomy. *Am Surg* 2014;80:377–81.
- Reznek M, Harter P, Krummel T. Virtual reality and simulation: training the future emergency physician. *Acad Emerg Med* 2002;9:78–87.
- Zirkle M, Blum R, Raemer DB, Healy G, Roberson DW. Teaching emergency airway management using medical simulation: a pilot program. *Laryngoscope* 2005;115:495–500.